## Spectro UV-VIS Double Beam PC Scanning Spectrophotometer

Model UVD-2960

## **Software Specifications**

Monoprocessor Built-in Application:

Photometric Measurement: Measuring transmittance or absorbance at the current wavelength together with K factor calculations.

Spectrum Scan: Carrying out scanning of transmittance or absorbance on the selected wavelength range together with peak-pick module.

Quantitative Determination: Regression of standard curves and direct determination concentration of samples.

New UVWin 6.0 PC Windows Application Software (RS-232 and USB Interface) to link Spectro to computer and printer:

Photometric Measurement: Measuring the photometric values at 1-10 wavelengths together with mathematical calculations according to entered quotations.

Spectrum Scan: Producing Wavelength scans within the operating parameters on samples together with powerful data handling facilities.

Quantitative Determination: Determination of unknown concentration with methods of 1-3 wavelength quantitation, together with fitting of calibration curve of 1st ~ 4th order.

Kinetics: Recording curves of changing photometric values of samples against timecourse at the selected wavelengths together with powerful data handling facilities.

Output: With the Windows clipboard, the measured data and graphics can be copied to other applications software for reports.

## **Technical Specifications**

Wavelength Range:	190 nm – 1100 nm	Baseline Stability:	0.0008Abs/h (2 hours warming up, 2nm Bandwidth, 500nd
Spectral Bandwidth:	0.5, 1.0, 2.0 and 5.0 nm	Slew Rate of Wavelength:	3600nm/min
Resolution:	0.1nm	DNA/RNA Measurement:	Results Printout: Printing of measured data
Straylight:	>2.10Abs (200nm)		by using any Printer with Parallel Port
Wavelength Accuracy:	±0.3 nm (with automatic wavelength correction).		connection available.
Wavelength Reproducibility:	±0.2 nm	Mainframe:	Compact and standalone spectrophotometer
Photometric System:	Double beam optical system		mainframe
Photometric Method:	Transmittance, absorbance, energy, concentration	Light Source:	Socket Deuterium Lamp and Socket
Photometric Range:	-0.3~3.0 Abs (0~200%T)		Tungsten Halogen Lamp
Photometric Accuracy:	±0.002Abs (0~0.5A), ±0.004Abs (0.5~1.0A),	Detector:	Double Beam
	±0.3%T (0-100%T)	Sample Chamber:	2 cell holder
Photometric Reproducibility:	±0.001Abs (0~0.5 A), ±0.002Abs (0.5~1.0A),	Display:	Liquid Crystal Display (LCD 320/240
	±0.15%T (0~100%T),		dot matrix)
Photometric Display:	-9999 9999	Keypad:	Touch soft keys.
Photometric Noise:	±0.001Abs (at 500nm, 2nm Spectral Bandwidth 0 Abs)	PC Interface:	PC Interface: RS-232, USB
Scanning Speed:	1400nm/min	Size:	22"x16"x10"
Baseline Flatness:	±0.0015Abs (2000 nm. ~ 1100 nm.)	Weight:	55 Lb
Voltage:	110V-240V		7